

INFECTION CONTROL DEPARTMENT

MEDICAL DEVICE PRE PURCHASE DECONTAMINATION QUESTIONNAIRE

Tenscare Ltd All TENS Units

Note: Failure to satisfactorily complete this document may influence the Trust's decision to purchase or use this device/equipment/system.

Important: this questionnaire is only for re-usable medical devices equipment and systems which cannot be decontaminated in a washer/disinfector (HTM 2030 refers) or in a porous load steam steriliser (HTM 2010 refers). For equipment which can be re-processed in either a washer/disinfector or steam steriliser please use questionnaire from the Trust's Central Sterile Supplies Department.

Cleaning and decontamination instructions must be provided for all equipment which is designed to be used by more than one patient. This includes non-invasive items which only have contact with patient's intact skin.

It is the supplier/manufacture's responsibility to recommend detergents and disinfectants which can safely be used to clean and decontaminate the equipment they supply to the Trust.

CHECKLIST

		Yes	No
1.	Is the Medical Device CE marked?	x	
2.	Are there clear and comprehensive instructions for cleaning/decontamination?	x	
3.	Do the instructions specify which detergents and disinfectants can be used?		x
4.	Can the product be cleaned with a neutral detergent?	x	
5.	Can the product be disinfected with a solution of sodium hypochlorite 1,000 ppm (approx equivalent to a 1:100 dilution of domestic bleach)?	x	
6.	Can the product be disinfected with an alcohol wipe (70% ethanol)?	x	
7.	Have you supplied a copy of the cleaning and decontamination instructions with this questionnaire?	x	

Declaration

I declare that the information provided is complete and accurate, and that should any information change, such changes will be brought to the attention of the Trust.

Completed by:	
Print name:	A S BROWN
Position:	SALES MANAGER
Telephone No.	01372 723434
Signature:	
For and on behalf of:	Tenscare Ltd
Date:	5/4/06

REFERENCES

Heath Technical Memorandum 2030
Health Technical Memorandum 2010